

EU-type examination certificate UK/0126/0012 Revision 14

Issued by:

NMO

Notified Body Number 0126

In accordance with the requirements of the Measuring Instruments Regulations 2016 (S.I. 2016 No. 1153) which implement, in the United Kingdom, Council Directive 2014/32/EU, this EU-type examination certificate has been issued to:

Elster Water Metering Limited
130 Camford Way
Sundon Park
Luton, Bedfordshire
LU3 3AN
United Kingdom

In respect of a family of cold water meters utilising a common, volumetric measuring element designated the V100/V110/V200/V210/V230 and having the following characteristics:

A nominal capacity of 36 revs/litre

A rated permanent flowrate Q_3 of 1.6m³/h or 2.5m³/h

The necessary data (principal characteristics, alterations, securing, functioning etc) for identification purposes and conditions (when applicable) are set out in the descriptive annex to this certificate.

This revision replaces previous versions of the certificate.

Issue date: 24 April 2018
Valid until 12 September 2027



G Stones
Technical Manager
For and on behalf of the Head of Technical Services

NMO | Stanton Avenue | Teddington | TW11 0JZ | United Kingdom

Tel +44 (0) 20 8943 7272 | Fax +44 (0) 20 8943 7270 | Web www.gov.uk/government/organisations/regulatory-delivery
NMO is part of the Office for Product Safety and Standards directorate within the Department for Business, Energy & Industrial Strategy



0135